



File Information

File Created Date: 09/30/19 23:23:13
 Note: All Times shown are based on UTC+0:00 and 24-Hour clock [MM/DD/YY HH:MM:SS]

Device Information

Device Type: TZ-TempU03 Firmware Version: 3.17
 ID: TZ1905220525
 Start Delay: 30 min Log Interval: 10 min
 Description: Temperature and Humidity Recorder

Logging Summary

First Point: 09/30/19 23:11:43 Max: 33.6°C(Temp)/51.9%RH(Humidity)
 Stop Time: 09/30/19 23:21:43 Min: 33.6°C(Temp)/51.0%RH(Humidity)
 Number of Points: 2 Average: 33.6°C(Temp)/51.4%RH(Humidity)
 Trip Length: 00d 00h 10m 00s MKT: 33.6°C
 Stop Mode: USB Stop

Marked Events

N/A

| Alarm Condition | Alarm Delay | First Point Time | Time of Violations | No. of Violations | Status |
|-----------------|-----------------|-------------------|--------------------|-------------------|--------|
| Over 8.0°C | 00d 00h 10m 00s | 09/30/19 23:11:43 | 00d 00h 10m 00s | 2 | Alarm |
| Below 2.0°C | 00d 00h 10m 00s | 00/00/00 00:00:00 | 00d 00h 00m 00s | 0 | OK |
| Over 80.0%RH | 00d 00h 10m 00s | 00/00/00 00:00:00 | 00d 00h 00m 00s | 0 | OK |
| Below 40.0%RH | 00d 00h 10m 00s | 00/00/00 00:00:00 | 00d 00h 00m 00s | 0 | OK |



