



File Information

File Created Date: 09/21/18 04:14:35
 Note: All Times shown are based on UTC+0 and 24-Hour clock [MM/DD/YY HH:MM:SS]

Device Information

Device Type: TZ-TempU03 Firmware Version: 3.05
 ID: TZ2018082224
 Start Delay: 30 min Log Interval: 10 min
 Description: Temperature and humidity Recording

Logging Summary

First Point: 09/20/18 05:15:35 Max: 23.0°C(Temp)/86.5%RH(Humidity)
 Stop Time: 09/21/18 04:05:35 Min: 1.7°C(Temp)/42.0%RH(Humidity)
 Number of Points: 138 Average: 6.3°C(Temp)/70.9%RH(Humidity)
 Trip Length: 00d 22h 50m 00s MKT: 11.3°C
 Stop Mode: USB Stop

Marked Events

N/A

Alarm Condition	Alarm Delay	First Point Time	Time of Violations	No. of Violations	Status
Over 8.0°C	00d 00h 10m 00s	09/20/18 23:25:35	00d 04h 40m 00s	29	Alarm
Below 2.0°C	00d 00h 10m 00s	09/20/18 06:55:35	00d 02h 50m 00s	18	Alarm
Over 60.0%RH	00d 00h 10m 00s	09/20/18 05:35:35	00d 18h 00m 00s	109	Alarm
Below 40.0%RH	00d 00h 10m 00s	00/00/00 00:00:00	00d 00h 00m 00s	0	OK



